

CLAIMS

- Sub A1
1. A method of protecting an x-ray analyzer comprising:
 applying a high voltage to a x-ray tube;
 obtaining discharge phenomenon that occurs in an x-ray generating
 system as a pulse array;
 counting a pulse number of the pulse array; and
 terminating the high voltage output to the x-ray tube when the pulse
 number reaches a predetermined value.
 2. An electric discharge detection circuit comprising:
 an X-ray tube;
 a power supply to generate a high voltage applied to the x-ray tube;
 an x-ray tube voltage detector to detect the high voltage applied to
 the x-ray tube;
 a differentiation circuit to differentiate a signal output from the
 x-ray tube voltage detector;
 a zero-crossing comparator to discriminate a polarity of an output
 signal from the differentiation circuit;
 a re-triggerable one-shot pulse generating circuit that generates a
 one-shot pulse at a fixed period, a pulse output from the zero-crossing comparator
 being a trigger of the one-shot pulse generating circuit;
 a counter, having a one-shot pulse output from the one-shot pulse
 generating circuit input as a operation enable signal, to count pulses output from
 the zero-crossing comparator during a period when operation is enabled;
 an x-ray cut-off circuit to transmit a command signal to the power
 supply to stop generation of high voltage when receiving a carry output from the
 counter; and
 a display to display occurrence of an electric discharge phenomenon
 upon receipt of the carry output from the counter.
 3. An electric discharge detection circuit comprising:

an X-ray tube;
a power supply to generate a high voltage applied to the x-ray tube;
an x-ray tube voltage detector to detect the high voltage applied to
the x-ray tube;

5

a discrimination circuit to discriminate between substantial changes
in a signal output from the x-ray tube voltage detector within a predetermined time
period;

an x-ray cut-off circuit to terminate generation of the high voltage
by the power supply when the number of substantial changes in the signal output
from the x-ray tube voltage detector within the predetermined time period exceeds
a preset amount; and

10

a display to display the preset amount has been exceeded.

11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60
61
62
63
64
65
66
67
68
69
70
71
72
73
74
75
76
77
78
79
80
81
82
83
84
85
86
87
88
89
90
91
92
93
94
95
96
97
98
99
100